


Acclaim columns overview

Columns for challenging separation needs

Broad range of HPLC and UHPLC columns



Broad range of HPLC and UHPLC columns for challenging separation needs

- **General-purpose RP**

Separate complex samples with high surface area columns (C18, etc)

- **Mixed-Mode**

Retain multiple types of analytes on a single column

- **Application-specific**

Unique columns for specific applications: surfactants, organic acids, pesticides, aminoglycosides, explosive residues

- **HILIC**

Designed for the separation of polar compounds

Contents

Acclaim column families at a glance	4
<hr/>	
Selecting an Acclaim column	5
<hr/>	
Acclaim reversed-phase, HILIC and SEC columns	8
<hr/>	
Acclaim Mixed-Mode HPLC columns	32
<hr/>	
Acclaim application-specific HPLC columns	42
<hr/>	
Acclaim column selection guide	59
<hr/>	
Ordering information	60
<hr/>	

Acclaim column families at a glance

Thermo Scientific™ Acclaim™ columns are designed for high-efficiency high performance liquid chromatography (HPLC) separations. Their unique functionalities set them apart for difficult and complex separations, often providing resolution of compounds not possible by other conventional HPLC columns.

Surface chemistries include standard functionalities such as Thermo Scientific™ Acclaim™ 120 C18, C8, Phenyl, Polar Advantage, polar embedded phases, HILIC, as well as specialty columns: Mixed-Mode functionalities (i.e. Thermo Scientific™ Acclaim™ Trinity P1 and P2 columns), application-specific phases (organic acids, surfactants, pesticides, pollutants, etc.) and size exclusion columns (SEC). Further, some Acclaim columns have been optimized to work under the very high pressure, ultra high performance liquid chromatography (UHPLC) requirements of Thermo Scientific™ Vanquish™ UHPLC instrument.

General purpose	Mixed-Mode	Application-specific
<ul style="list-style-type: none">• Start here for separations based on hydrophobic separations. These surface area Acclaim columns provides outstanding resolution for complicated matrix, large number of peaks, or purely need more retention.• Included are reversed-phase for hydrophobic interaction, HILIC for the retention or separation of hydrophilic compounds such catecholamines other polar compounds (specific).• SEC when you are looking for size exclusion separations, these unique columns have a polymer backbone eliminating secondary separation that can occur on silica media. They are robust with a wide pH range.	<ul style="list-style-type: none">• Turn to Mixed-Mode phases when you have a mix of ionic and hydrophobic compounds or hydrophilic compounds.• They are available in both dual and triple retention modes. Included are anionic/cationic/reversed-phase/HILIC functionalities. The bonding has been optimized for spacial distancing for greater resolution.• Over 10 years of experience for pharmaceutical APIs and counterions, including those required in standard operating procedures.	<ul style="list-style-type: none">• Take the guess-work out of your chromatography columns selection. These columns are designed and tested for specific applications.• Columns for food/beverage, consumer goods testing, and environmental testing. For those customers who run specific types of separations, these applications specific columns are the easy choice.• Wide range of column sizes to meet your chromatographic needs



Selecting an Acclaim column

Choosing the column chemistry is the first step in identifying the column of choice. This is key in achieving the resolution and selectivity for the desired separation. Particle size is important in determining the speed of a separation and column dimensions are important in optimizing resolution while also impacting solvent consumption and sensitivity requirements.

What are the goals of the separation?

- Critical peaks requiring resolution
- Speed of analysis
- Solvent consumption/minimal waste
- Sensitivity requirements

Choosing the column chemistry

When a chromatographer designs a separation, selectivity of the stationary phase is the first of many factors that must be considered. Selectivity is the result of the differing interactions between each analyte and the stationary phase. Determining the column with optimal selectivity is the essential starting point. This determination depends on the nature of the desired separation. The easiest first step in choosing a column is to identify whether there is a column designed specifically for your separation. This can be readily established if the column is named according to its application, as in the case of some of the Acclaim columns. Thus, if you are separating surfactants, the Thermo Scientific™ Acclaim™ Surfactant Plus column is the best first choice; likewise Thermo Scientific™ Acclaim™ Explosives column is used for Environmental Protection Agency (EPA) Method 8330, and the Thermo Scientific™ Acclaim™ Organic Acid column is for organic acid analysis.

If there is no specialty column for your application, then you will need some understanding of chemistry of your sample to choose the best column chemistry. For those cases where the nature of the sample is completely unknown, the best first choice is usually the C18 column because of its excellent peak efficiency, low silanol activity, and ability to separate many organic molecules. Other chemistries are available and can provide better chromatography based on the properties.

Type of bonding


The bonding chemistry of the column is defined by the functional groups attached to the particle substrate, as well as the substrate itself. In reversed-phase chromatography, the mode of separation is hydrophobic interaction, and the functional groups are either Thermo Scientific™ Acclaim™ C18, C8, C30, Polar Advantage (PA and PA2) or Phenyl. Other chemistries include HILIC, size exclusion chromatography and Mixed-Mode chromatography.

RP	Reversed-phase chromatography	Separation is based on hydrophobic interaction of the column surface chemistry and hydrophobicity of the molecules being separated. The stronger the hydrophobicity of the molecule – the more it will bind to the stationary phase
HILIC	Hydrophilic interaction liquid chromatography	Retains and separates polar, hydrophilic compounds
MM	Mixed-Mode chromatography	The stationary phase has more than one major form of retention, i.e., reversed-phase and ion-exchange
SEC	Size exclusion chromatography	Separation is based on molecular size

Choosing particle size and column format

Particle size affects peak efficiency, which in turn impacts resolution or the amount of space between peaks and the speed of the separation. Smaller particles sizes give higher efficiencies and higher resolution than larger ones. However, at the same flow rate, the backpressure will be greater using smaller particle sizes than larger ones. Therefore, smaller particle sizes are recommended for fast analyses, using shorter columns.

Acclaim columns are available packed with either 2.2 μm , 3 μm or 5 μm bonded silica. When using isocratic separation conditions, it can be expected that a column packed with 3 μm material will give about 1.7 times the separation efficiency as a column of identical proportions, packed with 5 μm material, but the backpressure will be almost three times as high on the 3 μm column at the same linear velocity. Similarly, for going from 3 μm to 2.2 μm , the efficiency increases 1.4 times, the resolution 1.2 times, and the pressure 1.9 times.



Column diameter

Select a column diameter according to your requirements for sample size, solvent consumption and sensitivity. Column diameter impacts sensitivity (narrow columns provide better sensitivity), loading capacity (wider is better for larger sample volumes and for purifying samples), solvent consumption (narrow columns use less solvent) and backpressure (narrow columns exhibit higher backpressure so you will need to decrease the flow rate). Scale the flow rate by the square of the column diameter to preserve backpressure and retention time. To gain the full benefits of smaller column diameters, your HPLC instrumentation needs to be optimized for the flow rate.

Column diameter selection

Column diameter	Application
2.1 mm internal diameter	High-sensitivity or limited volume samples. Use with low dead-volume instrumentation
3.0 mm internal diameter	Lower solvent consumption. Use with standard instrumentation
4.0-4.6 mm internal diameter	Standard formats. Use with standard instrumentation

Your system capabilities determine which size columns you are most likely to use. These suggested dimensions are for initial method development. Based on the general considerations, you might decide to use columns that are either longer or shorter these. Your optimized final method might use some other column or particle size.

System	Pressure	Particle	Column size	Flow
Classic LC	≤400 bar	3 – 5 μm	4.6 x 250 mm, 5 μm 4.6 x 150 mm, 3 μm	1 – 2 ml/min
Modern LC	≤600 bar	2 – 3 μm	3 x 150 mm, 3 μm 2.1 x 150 mm	0.3 – 0.5 ml/min
Thermo Scientific™ Vanquish™ UHPLC	≤1500 bar	≤ 2.2 μm	2.1 x 150 mm 2.1 x 250 mm	0.4 – 0.8 ml/min

Mobile phase considerations

The effect of mobile phase composition on column lifetime should be part of the consideration for designing a method. Use the highest practical quality of water, solvent, and buffer components; HPLC-grade material has low UV absorbance and is submicron-filtered by the manufacturer. To prevent fouling of your Acclaim column, use mobile phases that have been filtered through a filter that is 0.5 μm or smaller.



Acclaim reversed-phase,
HILIC and SEC columns

Acclaim reversed-phase, HILIC and SEC columns


Reversed-phase chromatography

Reversed-phase chromatography is the most widely used HPLC separation mechanism. Functional groups C18, C8, C30 and Phenyl are attached to the silica substrate either via a single attachment (monomeric bonding), or via multiple attachments (multidentate bonding) such as the Acclaim Polar Advantage II (PA2) column. Monomeric bonded phases provide higher column efficiencies, than polymeric phases, but polymeric phases are very stable under pH extremes.

These functional groups are hydrophobic and attract hydrophobic molecules in the sample. For example, Catechins are a class of polyphenolic flavanols widely distributed in plants, notably tea and cacao. As can be seen in this selectivity comparison, stationary phases with a mixture of polar and reversed-phase retention provide the best separation. The Acclaim PA2 column is resistant to the acidic conditions usually preferred for separating polyphenols and gives the best separation.

Acclaim C18, C30, PA, PA2 and Phenyl-1 columns, 3 µm; 3 x 150 mm	
HPLC system	Thermo Scientific™ UltiMate™ 3000 Rapid Separation (RS) system
Mobile phase	(A) Acetonitrile (B) 100 mM Formic acid + 20 mM ammonium formate (C) Water
Flow rate	0.60 mL/min
Injection volume	5 µL
Detection	UV, 254 nm, 5 Hz, 1s resp. time
Temperature	30 °C
Peaks	1. Theobromine 2. Gallic acid 3. Caffeine 4. Gallocatechin 5. Epigallocatechin 6. Catechin 7. Epicatechin 8. Epigallocatechin gallate 9. Gallocatechin gallate 10. Epicatechin gallate 11. Catechin gallate 12. 3,4,5-trihydroxy cinnamic acid

Gradient	-6	0	12	15
A%	5	5	50	50
B%	10	10	10	10
C%	85	85	40	40



Carbon load and end-capping

For reversed-phase chemistries, the percent (%) carbon is a rough guide to the capacity of the column. Given the same type of silica the higher carbon load is an indicator of surface coverage. Phases with higher carbon loads are more strongly hydrophobic, resulting in higher capacity, longer retention times, and often better resolution.

When the functional groups are attached to the silica particle, not all silanol groups on the surface of the silica particle are covered. Free silanol groups will interact with polar analytes, altering the retention times and often causing peak tailing for organic bases. To minimize these secondary interactions, the free silanol groups are end-capped. All Acclaim column packings are end-capped.

Choosing the pore size and surface area


Select a pore size appropriate for the molecular weight of your sample analytes. The pore size should be large enough to allow the sample molecules of interest to enter and pass through. If the pore size is too small, size exclusion effects can cause unwanted peak broadening.

MW < 15 kDa: use Acclaim 120, Acclaim PA or Acclaim PA2

MW < 150 kDa: use Acclaim 300


Generally speaking, the smaller the pore size, the greater the number of pores and the higher the surface area. The higher the surface area, the higher the capacity of the packing material. This means that the retention time of a given analyte will be shorter on a wide-pore material than on a narrow-pore one.

300 Å



Larger pores = Smaller surface area

120 Å



Smaller pores = Larger surface area

HILIC chromatography

HILIC is the recommended mode of separating polar compounds. These compounds are unretained under conventional reversed-phase conditions, but are retained using HILIC columns, without the requirement for ion-pair additives in the mobile phase. In this technique the stationary phase is polar and the aqueous portion of the mobile phase acts as the stronger solvent. Thus, the polar analytes can be retained and separated, even with 5-20% aqueous mobile phase.

Size exclusion chromatography (SEC)

Size exclusion chromatography (SEC) is a major mode of HPLC that employs porous particles in the column to separate molecules by virtue of their size in solution. Using isocratic conditions, the small molecules flow into the substrate pores and move slower than the large molecules, which are too big to move into the pores. The result is a separation of the molecules based on size.

Acclaim reversed-phase and general purpose HPLC columns

Columns	Description
Acclaim 120 C18	High-efficiency reversed-phase columns for small molecule and complex mixtures separations, including liquid chromatography mass spectrometry (LC-MS)
Acclaim 120 C8	Reversed-phase columns for small molecule with intermediate hydrophobic retention
Acclaim 300 C18	Wide-pore columns for LC-MS life science applications; protein and peptides
Acclaim Polar Advantage (PA)	Polar-embedded reversed-phase columns for separating polar and nonpolar analytes with 100% aqueous mobile phase compatibility
Acclaim Polar Advantage II (PA2)	Polar-embedded columns with enhanced hydrolytic stability pH 1.5 – 10, excellent for polar and non-polar analytes with 100% aqueous mobile phase compatibility and LC-MS
Acclaim C30	Very high hydrophobic retention; good for separating structurally related isomers
Acclaim Phenyl-1	Polar reversed-phase columns with high aromatic selectivity
Acclaim HILIC-10	Designed for unique selectivity of hydrophilic molecules
Acclaim SEC-1000 Acclaim SEC-300	Separation of water-soluble polymers using resin based columns

Acclaim reversed-phase and general purpose HPLC columns specifications

	Acclaim 120 C18	Acclaim 120 C8	Acclaim 300 C18	Acclaim PA	Acclaim PA2	Acclaim C30	Acclaim Phenyl-1	Acclaim HILIC-10
Bonded phase	Octadecylsilane	Octylsilane	Octadecylsilane	Embedded sulfonamide	Embedded amide	C30 alkylsilane	Phenyl	Proprietary
USP type	L1	L7	L1	L60	L60	L62	L11	-
End-capped	Yes	Yes	Yes	Yes	Yes	Proprietary	Yes	No
Starting material	Ultrapure silica							
Particle shape	Spherical							
Particle sizes	2.2 µm 3 µm 5 µm	2.2 µm 3 µm 5 µm	3 µm	2.2 µm 3 µm 5 µm	2.2 µm 3 µm 5 µm	3 µm 5 µm	3 µm 5 µm	3 µm 5 µm
Average pore diameter	120 Å	120 Å	120 Å	120 Å	120 Å	200 Å	120 Å	120 Å
Surface area	300 m ² /g	300 m ² /g	100 m ² /g	300 m ² /g	300 m ² /g	200 m ² /g	300 m ² /g	300 m ² /g
Total carbon content	18%	11%	7%	17%	17%	13%	-	-
pH range	2–8	2–8	2–8	2–8	1.5–10	2–8	2–8	2–8

Acclaim SEC specifications

Columns	Acclaim SEC-300	Acclaim SEC-1000
Substrate	Hydrophilic polymethacrylate resin	Hydrophilic polymethacrylate resin
Particle Shape	Spherical	Spherical
Particle Size	5 µm	7 µm
Pore Size	300 Å (multi-pore)	1000 Å (multi-pore)
Separation range for PEO*	100 – 50,000 Daltons	1,000 – 1,000,000 Daltons
Exclusion limit for PEO*	50,000 – 150,000 Daltons	3,000,000 – 7,500,000 Daltons
pH Range	2–12	2–12

*PEO = polyethylene oxides


Acclaim 120 C18 columns

High performance reversed-phase columns for the separation of small molecules


Thermo Scientific Acclaim 120 C18 series columns feature a densely bonded monolayer of octadecyldimethylsiloxane on a highly pure, spherical, silica substrate with 120 Å pore structure.

Acclaim 120 C18 columns are the classic reversed-phase columns. These columns are recommended for general-purpose reversed-phase applications where high surface coverage, low silanol activity, and excellent efficiency are required. Acclaim 120 C18 columns feature:

- Highly efficient, symmetrical peaks for difficult basic and chelating analytes
- Ultrapure silica substrate
- Optimized surface pretreatment, proprietary high density bonding process, and double endcapping
- Reliability designed into the manufacturing process and assured by thorough and appropriate testing
- High hydrophobicity and low polarity yield high selectivity for hydrophobic substances
- LC-MS compatible
- Wide range of applications in pharmaceutical, environmental, food testing, and product-quality testing for small molecules



A: Acclaim 120 C18 column, 5 μ m, 150 x 4.6 mm	
B, C: Acclaim UHPLC C18 column, 2.2 μ m, 50 x 2.1 mm	
Mobile phase	200 mM HOAc in 10% (v/v) MeOH
Flow rate	(A) 1.00 mL/min (B) 0.41 mL/min (C) 0.82 mL/min
Injection volume	(A) 10 μ L (B) 1.2 μ L (C) 1.2 μ L
Detection	UV, 254 nm (A) 1 Hz data rate (B) 5 Hz data rate (C) 10 Hz data rate
Temperature	20 $^{\circ}$ C
Sample	Commercial vanilla extract in 40% ethanol, filtered
Reference	AOAC official method 990.25
Analytes	1. p-Hydroxybenzoic acid 2. p-Hydroxybenzaldehyde 3. Vanillic acid 4. Vanillin



LC-MS compatibility


Acclaim 120 C18 columns are LC-MS compatible with very low bleed. Out of the box, this column is ready to use after only a few minutes of conditioning with solvent. The same low bleed is attained over the entire pH range of 2–8.

Reproducible results of Peptide Map overlay of five chromatograms for the separation of digested infliximab

2 x Acclaim Vanquish C18 column, 2.2 μ m, 2.1 x 250 mm


Mobile phase	(A) Water + 0.1% formic acid (B) Water/acetonitrile (10:90 v/v) + 0.1% formic acid
Flow rate	0.4 mL/min
Detection	UV, 214 nm
Temperature	60 $^{\circ}$ C, forced air

Time (min)	0	40	40.1	43
A%	99	55	99	99
B%	1	45	1	1




Isocratic resolution of antihistamines and their impurities on Acclaim 120 C18 column

Acclaim 120 C18 column, 5 μ m, 4.6 \times 150 mm	
Mobile phase	(A) 50 mM sodium acetate (B) Methanol
Isocratic	(A) 20% (B) 80%
Detection	UV, 249 nm
Temperature	25 $^{\circ}$ C
Peaks	1. Thenyldiamine HCl 2. Phenothiazine 3. Promethazine HCl 4. Pyrrobutamine phosphate



Separation of basic drugs on Acclaim 120 C18 column in various concentrations

Acclaim 120 C18 column, 5 μ m, 4.6 \times 150 mm	
Mobile phase	80/20 methanol/30 mM phosphate, pH 6
Flow rate	1 mL/min
Injection volume	5 μ L
Detection	UV, 220 nm
Temperature	30 $^{\circ}$ C
Amitriptyline mass	Trace A: 1200 ng Trace B: 400 ng (normalized peak height) Trace C: 94 ng (normalized peak height)
Peaks	1. Uracil 2. Propranolol 3. Toluene 4. Doxepin 5. Amitriptyline




Note: The high performance of this column is maintained as the sample is diluted

Nitrofurantoin antibiotic residues in animal feed on Acclaim 120 C18 column



Acclaim 120 C18 column, 5µm, 4.6 × 150 mm	
Mobile phase	Acetonitrile:10 mM ammonium acetate pH 5.0, 20:80 v/v; isocratic
Flow rate	1.0 mL/min
Injection volume	20 µL
Detection	UV, 365 nm
Temperature	30 °C
Sample	<ol style="list-style-type: none"> 3.0 g guinea pig feed in a 50 mL centrifuge tube Add 9 mL water and let stand 5 min Add 21 mL methanol:acetonitrile 1:1 and extract for 30 min. Pass through cleanup cartridge containing 1.7 g of neutral alumina; discard first 1.7 mL, retain next 3.5 mL
Peaks	<ol style="list-style-type: none"> Nitrofurazone Nitrofurantoin Furazolidone Furaltadone

Reference: R.J. McCracken, D.G. Kennedy; J. Chromatogr. A, 1997, 771, 349-354.



Bitter principles in beer on Acclaim 120 C18 column

Acclaim 120 C18 column, 5µm, 4.6 × 150 mm	
Mobile phase	75% methanol, 24% water, 1% phosphoric acid (v/v/v)
Flow rate	1.2 mL/min
Injection volume	25 µL
Detection	UV, 270 nm
Temperature	TCC-100 thermostat, 35 °C
Sample	<p>Extraction per ASBC</p> <ul style="list-style-type: none"> 10 mL beer + 1 mL 3N HCl + 50 µL 1-octanol + 20 mL iso-octane Shake vigorously Centrifuge to separate phases Bitterness units = 50 × A275 <p>For HPLC</p> <ul style="list-style-type: none"> Evaporate dry and reconstitute in mobile phase <p><i>Obtained from Am. Soc. Brewing chemists</i></p>
Peaks	<ol style="list-style-type: none"> Isocohumulone Mixed isohumulone congeners Isoadhumulone




Acclaim 120 C8 columns

High performance reversed-phase columns with intermediate hydrophobic retention

Thermo Scientific Acclaim 120 C8 series columns feature a densely bonded monolayer of octyldimethylsiloxane on a highly pure, spherical silica substrate with a 120 Å pore structure.


Acclaim 120 C8 columns employ the same bonding chemistry and substrate as C18, and therefore features the same high standards of efficiency, coverage, and silanol activity. Acclaim 120 C8 column features:

- Similar selectivity to C18 columns, but with reduced retention
- Highly efficient, symmetrical peaks with difficult basic and chelating analytes
- Ultrapure silica substrate
- Optimized surface pretreatment, proprietary high-density bonding process, and vigorous endcapping
- Reliability designed into the manufacturing process and assured by thorough and appropriate testing
- Less hydrophobic, less retentive than C18
- LC-MS compatible
- Excellent performance for basic pharmaceuticals and environmental samples




Triclosan in toothpaste

Acclaim RSLC C8 column, 2.2 μm, 50 x 2.1 mm	
Mobile phase	Isocratic, 15% buffer, (2 mM Ammonium acetate pH 5), 85% methanol (v/v)
Flow rate	0.2 mL/min
Injection volume	1.0 μL
Detection	Diode array detector, 281 nm, 10 Hz, 0.1 s resp. time and spectra 200–400 nm
Temperature	50 °C
Sample	Toothpaste containing 0.3% triclosan
Analytes	1. Saccharin 2. Triclosan




Isocratic separation of six cardiac antiarrhythmic drugs (beta-blockers) on Acclaim 120 C8 column

Acclaim 120 C8 columns, 3 μm, 4.6 × 150 mm	
Mobile phase	51/49 w/w MeOH/ 25 mM phosphate, pH 7.0
Flow rate	1.0 mL/min
Injection volume	5 μL
Detection	UV, 214 nm
Temperature	40 °C
Analytes	1. Maleic acid (-) 2. Acebutolol (50 μg/mL) 3. Metoprolol (50) 4. Timolol (100) 5. Oxprenolol (50) 6, 7. Labetalol diastereomers (50) 8. Propranolol (20)



Hydrocortisone in skin ointment

Acclaim 120 C8 column, 3 μm, 4.6 × 150 mm	
Mobile phase	Acetonitrile:water 49:51
Flow rate	1.0 mL/min
Injection volume	1 μL
Detection	UV, 245nm
Temperature	30 °C
Sample	<ul style="list-style-type: none"> Disperse 100 mg of ointment in 7.0 mL denatured ethanol Mix in 3.0 mL hexane then 1.0 mL water Centrifuge to separate emulsion Filter ethanol layer
Peaks	1. Hydrocortisone 1.05% (label 1.0%) 2. Methyl paraben 0.32% 3. Propyl paraben 0.21%



Acclaim 300 C18 columns

High-resolution reversed-phase separation of proteins and peptides


Thermo Scientific Acclaim 300 C18 series columns feature a densely bonded monolayer of octadecyldimethylsiloxane on a highly pure, spherical silica substrate with a wider, 300 Å pore structure.

Acclaim 300 series columns are designed for small protein (up to 150 kDa) and peptide separations. Acclaim 300 columns are also useful for general-purpose, reversed-phase chromatography of small molecules.

Acclaim 300 C18 columns feature:


- Technology designed for high-resolution peptide and protein separations
- High efficiency 3 µm spherical silica substrate
- The same high performance bonding chemistry as the Acclaim 120 series, but using a silica substrate with larger 300 Å pores and lower surface area
- Minimal secondary interactions for repeatable results day-to-day and column-to-column
- LC-MS compatible

The unique bonding chemistry results in a high-density, highly uniform phase coverage with extensive endcapping. The use of a 3 µm silica particle accelerates the diffusion of the mobile phase into the stationary phase, resulting in fast, high-resolution separations. Compared to 5 µm column packings, a given separation can be achieved in a shorter run time by increasing the flow rate of the mobile phase and running shallower gradients on shorter columns.




Budesonide and related substances

Acclaim 300 C18 column, 3 µm, 150 x 4.6 mm	
Mobile phase	(A) Acetonitrile:ethanol 15:1 (B) 0.1% phosphoric acid isocratic 66% B
Flow rate	1.0 mL/min
Injection volume	15 µL
Detection	UV, 240nm
Temperature	30 °C
Sample	Budesonide, 500 µg/mL after three days
Analytes	7, 8. Budesonide epimers, 99%
Reference:	Hou S, Hindle M, Byron PR; J. Pharm. Biomed. Anal. 2001 24:371-80.




Tryptic Peptide Map of bovine serum albumin on Acclaim 300 C18 column

Acclaim 300 C18 column, 3 μm, 4.6 × 50 mm	
Mobile phase	(A) 95/5/0.1 H ₂ O/acetonitrile/TFA (B) 5/95/0.1 H ₂ O/acetonitrile/TFA
Flow rate	1.0 mL/min
Injection volume	40 μL
Detection	UV, 214 nm
Gradient	5% B to 40% B in 35 min



Comparison of acid additives for peptide analysis on Acclaim 300 C18 column


Acclaim 300 C18 column, 3 μm, 4.6 × 150 mm	
Mobile phase	(A) 70% acetonitrile v/v (B) Water (C) Acid stock solution
Flow rate	1.0 mL/min
Injection volume	5 μL
Detection	UVD 340U; UV, 214 nm
Temperature	TCC100 thermostat at 30 °C
Acid additive	A. Phosphoric acid, 0.15% B. Trifluoroacetic acid, 0.1% C. Formic acid, 0.1%
Sample	Sigma H2016 Peptide Mix
Peaks	1. Gly-Tyr 2. Val-Tyr-Val 3. Met-Enkephalin 4. Angiotensin-II 5. Leu-Enkephalin



Gradient	-8	0	16
A%	7	7	75
B%	68	68	0
C%	25	25	25

Protein mixture by reversed-phase HPLC on Acclaim 300 C18 column

Acclaim 300 C18 column, 3 μm, 4.6 × 150 mm	
Mobile phase	(A) 5 mM MSA in 95/5 H ₂ O/acetonitrile (B) 5 mM MSA in 5/95 H ₂ O/acetonitrile
Injection volume	25 μL
Detection	UV, 214 nm
Flow rate	1.0 mL/min
Peaks	1. RNase A 2. Impurity in carbonic anhydrase 3. Impurity in lysozyme 4. Lysozyme 5. Impurity in carbonic anhydrase 6. Myoglobin 7. Carbonic anhydrase




Gradient	0	3	18	23
A%	100	100	0	0
B%	0	0	100	100

Acclaim C30 columns

Columns for separating structurally related isomers


The Thermo Scientific Acclaim C30 column is designed to provide high shape selectivity for separating hydrophobic structural related isomers and unique selectivity complementary to other reversed-phase columns (e.g., C18).

- High shape selectivity
- Unique selectivity complementary to other reversed-phase columns
- Compatibility with highly aqueous mobile phase
- High-quality: low column bleed, high efficiency and rugged packing



Carotenoids in vegetables


Acclaim C30 column, 5 μm, 150 x 4.6 mm					
LC system	UltiMate 3000 RS				
Mobile phase	(A) Acetonitrile (B) Methanol:Ethyl acetate 1:1 (v/v) (C) 200 mM acetic acid in water				
Injection volume	28 μL				
Detection	Diode Array; VIS 450 nm, spectra 260-800 nm				
Flow rate	1.50 mL/min				
Temperature	30 °C				
Sample	A. Carrot extract in acetone B. Spinach extract in acetone C. Maize extract in acetone				
Peaks	1. Lutein 2. Zeaxanthin 3. Chlorophyll-b 4. alpha-Cryptoxanthin 5. beta-Cryptoxanthin 6. Chlorophyll-a 7. alpha-Carotene 8. beta-Carotene				
Time	-5	0	2	15	25
A%	85.0	85.0	85.0	65.0	65.0
B%	14.5	14.5	14.5	34.5	34.5
C%	0.5	0.5	0.5	0.5	0.5



Reference: "HarvetsPlus Handbook for Carotenoid Analysis", D.B. Rodriguez- Amaya and M. Kimura, International Food Policy Research Institute, 2004.

Glucocorticosteroids

Acclaim C30 column, 3 μm, 150 x 3.0 mm	
LC system	UltiMate 3000 RS system
Mobile phase	Methanol:tetrahydrofuran:water 3:25:72(v/v)
Injection volume	2 μL
Detection	Diode Array, UV, 240 nm
Flow rate	0.50 mL/min
Temperature	30 °C
Samples	A. Carrot extract in acetone B. Spinach extract in acetone C. Maize extract in acetone
Peaks (50 μg/mL)	1. Prednisone 2. Cortisone 3. Prednisolone 4. Hydrocortisone 5. Dexamethasone 6. 6-Methylprednisolone 7. Corticosterone 8. 11-Deoxyhydrocortisone




Reference: McWhinney B C, Ward G, Hickman P E; *Clin. Chem*, 1996 , 42:979-981.

Analysis of cooking oils

Acclaim C30 column, 5 μ m, 150 x 4.6 mm	
LC system	UltiMate 3000 RS system
Mobile phase	Acetonitrile (MeCN)/Iso-propanol (IPA)/ Ammonium Acetate (0.1 M, pH5.0) (Buffer)
Injection volume	2 μ L
Detection	Corona <i>ultra</i> (Gain = 100 pA; Filter = medium; Neb. temp = 25 $^{\circ}$ C)
Flow rate	1.0 mL/min
Temperature	40 $^{\circ}$ C
Sample	Peanut, olive, or wheat germ oil (5 mg/mL in iso-propanol)


Time (min)	MeCN	IPA	Buffer
-15	90	5	5
0	90	5	5
0.1	90	5	5
60	0	95	5
80	0	95	5



Profile of egg lecithin

Acclaim C30 column, 5 μ m, 150 x 4.6 mm	
LC system	UltiMate 3000 RS system
Mobile phase	Acetonitrile (MeCN)/Iso-propanol (IPA)/ Ammonium Acetate (0.1 M, pH5.0) (Buffer)
Injection volume	2 μ L
Detection	Corona <i>ultra</i> (Gain = 100 pA; Filter = medium; Neb. Ttemp = 25 $^{\circ}$ C)
Flow rate	1.0 mL/min
Temperature	40 $^{\circ}$ C
Sample	Peanut, olive, or wheat germ oil (5 mg/mL in iso-propanol)

Time (min)	MeCN	IPA	Buffer
-15	70	0	30
0	70	0	30
0.1	70	0	30
10	90	0	10
35	10	80	10
50	0	95	5
60	0	95	5




Acclaim Polar Advantage (PA) columns

Novel polar-embedded reversed-phase columns with unique selectivity

The Thermo Scientific Acclaim PA column has a patented surface chemistry that renders it compatible with solvent-free mobile phases. The ether and sulfonamide linkages are more hydrolytically stable than the amides used in many polar-embedded phases. The synthesis procedure minimizes both residual silanols and amines, thus making the Acclaim PA column suitable for acidic, basic, or neutral analytes.

The Acclaim PA column benefits include:

- Compatibility with solvent-free applications without any compromise to performance for acids and bases
- Novel polar-embedded surface layer
- Ability to work with 0–100% aqueous or 0–100% organic solvent mobile phases
- Resolves hydrophilic compounds
- High selectivity for hydrophobic compounds
- Different selectivity than C18 makes PA useful as a confirmation column
- Wide range of applications in pharmaceutical, environmental, food testing, and product-quality testing
- LC-MS compatible



EPA method 604 Phenols separation

Acclaim RSLC Polar Advantage column, 2.2 μm, 50 x 3.0 mm

Mobile phase (A) 10mM formic acid + 10 mM ammonium formate, pH 3.75 ± 0.05
(B) Acetonitrile

Flow rate 1.25 mL/min


Injection volume 0.5 μL

Detection UV, 280nm, 10Hz, 0.5 s resp. time

Temperature 30 °C

Peaks

1. Phenol
2. 2,4-Dinitrophenol
3. 4-Nitrophenol
4. 2-Chlorophenol
5. 2-Nitrophenol
6. 2,4-Dimethylphenol
7. 4,6-Dinitro-2-methylphenol
8. 4-Chloro-3-methylphenol
9. 2,4-Dichlorophenol
10. 2,4,6-Trichlorophenol
11. Pentachlorophenol



Gradient	-1.5	0.0	0.3	2.6	3.0
A%	70	70	70	10	10
B%	30	30	30	90	90

Resistance to dewetting example

The surface of a conventional C18 phase is very hydrophobic. When hydrophobic surfaces are in contact with highly aqueous mobile phases, the partial pressure of dissolved gases can expel the mobile phase from the pores of the stationary phase. This process is called dewetting and it adversely affects chromatographic performance. By design, the mildly hydrophilic surface of the Acclaim PA column remains in contact with aqueous-only mobile phases, negating the problem of dewetting.



While the onset of dewetting is somewhat unpredictable, stopping the flow of mobile phase through the column can initiate the process. The accompanying figure shows the effect of repeatedly stopping the flow through a C18 and a PA column. The C18 column dewets in a single cycle, but the PA column remains wetted through many cycles.

Resistance to dewetting

Acclaim PA column, 5 μ m, 4.6 \times 150 mm	
Mobile phase	2.5 mM methanesulfonic acid, pH 2.6
Flow rate	1 mL/min
Injection volume	5 μ L
Detection	UV, 254 nm
Temperature	30 $^{\circ}$ C
Peaks	1. Cytosine 2. Uracil 3. Thymine


Each cycle consists of two steps:

1. Equilibrate columns for 20 min before testing for 10 min.
2. Stop flow for 30 min before next cycle begins.



Separation of five antidepressants on Acclaim PA column

Acclaim PA column, 5 μ m, 4.6 \times 150 mm	
Mobile phase	80/20 v/v MeOH/30 mM phosphate, pH 6.0
Flow rate	1 mL/min
Injection volume	5 μ L
Detection	UV, 220 nm
Temperature	30 $^{\circ}$ C
Peaks	1. Protriptyline 50 2. Nortriptyline 25 3. Doxepin 50 4. Imipramine 40 5. Amitriptyline 50




Acclaim Polar Advantage II (PA2) columns

Complementary selectivity and enhanced hydrolytic stability

Thermo Scientific Acclaim Polar Advantage II (PA2) columns, like the Acclaim PA2 column, is a high-efficiency, silica-based, reversed-phase column with a polar enhanced stationary phase for operation over a wider range of chromatographic conditions than is possible with conventional reversed-phase stationary phases.

Acclaim PA2 column is an amide polar-embedded phase, with all the advantages of conventional polar-embedded phases, but with enhanced hydrolytic stability at both low and high pH (pH 1.5–10). The Acclaim PA2 column provides selectivity that is complementary to conventional C18 columns, and to our Acclaim PA column. This column is fully compatible with 100% aqueous mobile phases and provides symmetrical peaks for both polar and non-polar analytes.



- Ability to separate polar and non-polar compounds
- Exceptional hydrolytic stability (pH 1.5–10)
- High polarity for complementary selectivity to C18 columns
- Compatibility with 0–100% aqueous or 0–100% organic solvent mobile phases
- Good peak shapes for both acidic and basic compounds
- High column efficiency
- Broad range of applications in pharmaceutical, environmental, food testing and product-quality testing



Turmeric separation


Acclaim 120 C18 column, 2.2 μm, 100 x 2.1 mm
Acclaim PA2 column, 2.2 μm, 100 x 2.1 mm

Mobile phase	(A) 15mM H ₃ PO ₄ (B) Methanol
Flow rate	0.41 mL/min
Isocratic	C18: 70% B (v/v) PA2: 80% B (v/v)
Detection	UV, 428 nm
Sample	Turmeric extract
Temperature	30 °C
Analytes	1. Curcumin 2. Demethoxycurcumin 3. Bis-demethoxycurcumin




Beta blockers on Acclaim PA2 columns

Acclaim PA2 column, 5 μ m, 4.6 \times 150 mm									
Mobile phase	MeOH/0.2% NH ₄ OH, pH10, v/v, 60/40								
Flow rate	1.0 mL/min								
Injection volume	5 μ L								
Detection	UV, 210 nm								
Temperature	30 °C								
Peaks (40 ppm each)	<table border="0"> <tr> <td>1. Maleate</td> <td>5. Acebutolol</td> </tr> <tr> <td>2. Labetalol</td> <td>6. Metoprolol</td> </tr> <tr> <td>3. Metaraminol</td> <td>7. Timolol</td> </tr> <tr> <td>4. Atenolol</td> <td>8. Oxprenolol</td> </tr> </table>	1. Maleate	5. Acebutolol	2. Labetalol	6. Metoprolol	3. Metaraminol	7. Timolol	4. Atenolol	8. Oxprenolol
1. Maleate	5. Acebutolol								
2. Labetalol	6. Metoprolol								
3. Metaraminol	7. Timolol								
4. Atenolol	8. Oxprenolol								




Catecholamines in urine on Acclaim PA2 columns

Acclaim PA2 column, 3 μ m, 2.1 \times 150 mm													
Mobile phase	Buffer (11.98 g citric acid, 3.53 g NaOAc, 37.2 mg EDTA, 10 mL of 0.1 M methanesulfonic acid in 1 L D.I. water)/MeOH v/v 90/10												
Flow rate	0.2 mL/min												
Injection volume	2 μ L												
Detection	DC amperometry [GC electrode, 800 mV]												
Temperature	30 °C												
Peaks (1 nM each)	<table border="0"> <tr> <td>1. 4-Hydroxy-3-methoxymandelic acid</td> <td>5. 3,4-dihydroxybenzylamine</td> </tr> <tr> <td>2. 4-Hydroxy-3-methoxyphenylglycol</td> <td>6. Normetanephrine</td> </tr> <tr> <td>3. Norepinephrine</td> <td>7. Metanephrine</td> </tr> <tr> <td>4. Epinephrine</td> <td>8. Dopamine</td> </tr> <tr> <td>9. 4-Hydroxy-3-methoxyphenylacetic acid</td> <td></td> </tr> <tr> <td>10. 5-Hydroxyindoleacetic acid</td> <td></td> </tr> </table>	1. 4-Hydroxy-3-methoxymandelic acid	5. 3,4-dihydroxybenzylamine	2. 4-Hydroxy-3-methoxyphenylglycol	6. Normetanephrine	3. Norepinephrine	7. Metanephrine	4. Epinephrine	8. Dopamine	9. 4-Hydroxy-3-methoxyphenylacetic acid		10. 5-Hydroxyindoleacetic acid	
1. 4-Hydroxy-3-methoxymandelic acid	5. 3,4-dihydroxybenzylamine												
2. 4-Hydroxy-3-methoxyphenylglycol	6. Normetanephrine												
3. Norepinephrine	7. Metanephrine												
4. Epinephrine	8. Dopamine												
9. 4-Hydroxy-3-methoxyphenylacetic acid													
10. 5-Hydroxyindoleacetic acid													



Separation of water-soluble vitamins on Acclaim PA2 columns

Acclaim RSLC Polar Advantage PA2 column, 2.2 μ m, 150 \times 2.1 mm											
Mobile phase	(A) 25 mM potassium dihydrogen phosphate, pH 3.8 (adjusted with phosphoric acid) (B) 70:30 acetonitrile/25 mM potassium dihydrogen phosphate, pH 3.8										
Flow rate	0.4 mL/min										
Injection volume	2 μ L										
Detection	UV, 210 nm										
Temperature	25 °C (with passive pre-heater)										
Peaks	<table border="0"> <tr> <td>1. Ascorbic acid</td> <td>6. Pantothenic acid</td> </tr> <tr> <td>2. Thiamine</td> <td>7. Folic acid</td> </tr> <tr> <td>3. Pyridoxal</td> <td>8. Cyanocobalamin</td> </tr> <tr> <td>4. Pyridoxin</td> <td>9. Riboflavin</td> </tr> <tr> <td>5. Nicotinamide</td> <td></td> </tr> </table>	1. Ascorbic acid	6. Pantothenic acid	2. Thiamine	7. Folic acid	3. Pyridoxal	8. Cyanocobalamin	4. Pyridoxin	9. Riboflavin	5. Nicotinamide	
1. Ascorbic acid	6. Pantothenic acid										
2. Thiamine	7. Folic acid										
3. Pyridoxal	8. Cyanocobalamin										
4. Pyridoxin	9. Riboflavin										
5. Nicotinamide											




Time	0	5	7	10
B%	0	36	10	0

Acclaim Phenyl-1 column

A unique reversed-phase column with high aromatic selectivity

Thermo Scientific Acclaim Phenyl-1 columns are designed to provide unique selectivity distinguished from other reversed-phase HPLC columns, resulting in superior separations for analytes that cannot be resolved well on typical alkyl phases (C18 and C8) or other phenyl-type columns.

- High aromatic selectivity
- High hydrophobic retention
- Unique and complementary selectivity compared to any other phenyl type column
- Compatibility with highly aqueous mobile phase
- High efficiency and rugged packing




Acclaim Phenyl-1 column has a higher pi-pi interaction than other phenyl phases and provides unique selectivity for aromatic compounds while maintaining sufficient hydrophobic interaction and aqueous compatibility for superior chromatographic performance.

Acclaim Phenyl-1 column can be used in a wide range of applications in pharmaceutical, environmental, food testing and product-quality testing. This column is ideally suited for the analysis of aromatic analytes; some examples include glucocorticosteroids, estrogens, fat-soluble vitamins and phospholipids.


Separation of fat-soluble vitamins

Acclaim Phenyl-1 column, 3 μm, 150 x 3.0 mm	
Mobile phase	Methanol/water v/v 90/10
Flow rate	0.5 mL/min
Injection volume	2 μL
Detection	UV, 220nm
Temperature	30 °C
Peaks (100 ppm each)	1. Retinol acetate (vitamin A acetate) 2. Vitamin D2 3. Vitamin D3 4. delta-Tocopherol 5. gamma-Tocopherol 6. alpha-Tocopherol (vitamin E) 7. Impurity (unknown) 8. Vitamin E acetate 9. Vitamin K2 10. Vitamin K1




Separation of glucocorticosteroids

Acclaim Phenyl-1 column, 3 μ m, 3.0 \times 250 mm	
Mobile phase	Methanol/water v/v 46/54
Flow rate	0.5 mL/min
Injection volume	5 μ L
Detection	UV, 254 nm
Temperature	40 $^{\circ}$ C
Peaks (50 ppm each)	1. Prednisone 2. Cortisone 3. Prednisolone 4. Hydrocortisone 5. Dexamethasone 6. 6-Methylprednisolone 7. Corticosterone 8. Deoxyhydrocortisone




Analysis of soybean and egg lecithin

Acclaim Phenyl-1 column, 3 μ m, 3.0 \times 150 mm	
Mobile phase	Acetonitrile/2-propanol/ammonium acetate (100 mM, pH5.0) v/v/v 45/45/10
Flow rate	0.45 mL/min
Injection volume	2 μ L
Detection	Corona ultra (Gain = 100 pA; Filter = low; Nebulizer temp. = 25 $^{\circ}$ C)
Temperature	25 $^{\circ}$ C
Samples	1. Soybean Lecithin (1 mg/mL in 2-propanol) 2. Egg lecithin (2.5 mg/mL in 2-propanol)



Phospholipids in soybean lecithin

Acclaim Phenyl-1 column, 3 μ m, 3.0 \times 150 mm	
Mobile phase	Acetonitrile/2-propanol/ammonium acetate (100 mM, pH 5.0) v/v/v 70/10/20
Flow rate	0.45 mL/min
Injection volume	5 μ L
Detection	Corona ultra (Gain = 100 pA; Filter = low; Nebulizer temp. = 25 $^{\circ}$ C)
Temperature	25 $^{\circ}$ C
Sample	Soybean lecithin (1 mg/mL in 2-propanol)



Acclaim HILIC-10 column


Designed with unique selectivity for hydrophilic molecules

The Thermo Scientific Acclaim HILIC-10 column is designed for separating highly hydrophilic molecules by hydrophilic interaction liquid chromatography (HILIC). This column is based on high-purity spherical porous silica covalently modified with a proprietary hydrophilic layer.

HILIC is a complementary technique to reversed-phase liquid chromatography (RPLC) with several benefits. Polar analytes that cannot be retained using RP columns can be retained and separated using the Acclaim HILIC-10 column. The advantage of the HILIC phases is that they allow the use of 5–20% aqueous mobile phase, while maintaining affinity for polar analytes.


The Acclaim HILIC-10 column is ideally suited for analysis of polar analytes as demonstrated its use in a wide variety of applications including pharmaceuticals, metabolites, fat-soluble vitamins, oils, industrial applications, etc.

- Retains highly polar molecules that are not retained by reversed-phase chromatography
- Unique selectivity, complementary to reversed-phase columns
- Hydrolytically stable
- Rugged column packing
- Broad application range




Glycerides

Acclaim HILIC-10 column, 3 μm, 150 x 3.0 mm	
Mobile phase	(A) Heptane (B) 2-Propanol/acetic acid 99.5:0.5
Flow rate	0.50 mL/min
Injection volume	4 μL
Detection	Corona ultra, nebulizer 15 °C
Temperature	25 °C
Analytes	1. Tristearin 5. Distearin isomer 2 2. Trilaurin 6. Dilaurin isomer 2 3. Distearin isomer 1 7. Monostearin 4. Dilaurin isomer 1 8. Monolaurin




Separation of hydrophilic pharmaceuticals

Acclaim HILIC-10 column, 3 μm, 4.6 x 150 mm	
Mobile phase	90/10 v/v CH ₃ CN/10 mM (total) NH ₄ OAc, pH 5
Flow rate	1 mL/min
Injection volume	2 μL
Detection	UV, 230 nm
Temperature	30 °C
Analytes	1. Acetaminophen 0.1 mg/mL 2. Salicylic acid 0.1 3. Aspirin 0.2 4. Penicillin G 0.1 5. Metformin 0.1




Separation of acrylic acid and oligomers

Acclaim HILIC-10 column, 3 μ m, 4.6 \times 150 mm	
Mobile phase	90/10 v/v CH ₃ CN/10 mM (total) NH ₄ OAc, pH 5
Flow rate	1 mL/min
Injection volume	2 μ L
Detection	Corona CAD <i>ultra</i>
Temperature	30 $^{\circ}$ C
Analytes	1. Cyanuric acid 0.2 mg/mL 2. Melamine 0.2




Separation of Good's buffer salts

Acclaim HILIC-10 column, 3 μ m, 4.6 \times 150 mm	
Mobile phase	85/15 v/v CH ₃ CN/10 mM (total) NH ₄ OAc, pH 5
Flow rate	1 mL/min
Injection volume	10 μ L
Detection	Corona CAD <i>ultra</i>
Temperature	30 $^{\circ}$ C
Analytes (0.1 mg/mL in mobile phase)	1. TAPS 2. CHES 3. MOPS 4. TES 5. HEPES



Separation of mono-, di-, and triglycerides

Acclaim HILIC-10 column, 3 μ m, 3.0 \times 150 mm	
Mobile phase	(A) Heptane (B) 2-Propanol-acetic acid 99.5:0.5
Flow rate	0.50 mL/min
Injection volume	4 μ L
Detection	Corona <i>ultra</i> , nebulizer 15 $^{\circ}$ C
Temperature	25 $^{\circ}$ C
Analytes	1. Tristearin 2. Trilaurin 3. Distearin isomer 1 4. Dilaurin isomer 1 5. Distearin isomer 2 6. Dilaurin isomer 2 7. Monostearin 8. Monolaurin




Time	-6.0	0.0	0.5	4.0	10.0
A%	99	99	96	87	87
B%	1	1	4	13	13

Acclaim size exclusion chromatography (SEC)

High performance SEC columns for analysis of water soluble polymers


Thermo Scientific Acclaim SEC-300 and SEC-1000 columns are a family of resin based, high-performance size exclusion chromatography columns specifically designed for the separation of water soluble polymers and oligomers.

- Proprietary mono-dispersed multi-pore hydrophilic resin: no inflection points in calibration curve
- SEC-300 columns calibrated from 100 to 50,000 Daltons
- SEC-1000 columns calibrated from 1,000 to 1,000,000 Daltons
- Availability of small particle sizes packed in 300 x 4.6 mm dimension allows for high-resolution analysis at reduced solvent consumption
- Stable surface bonding with low column bleed and compatibility with UV, RI, MS, ELSD and Thermo Scientific™ Dionex™ Corona™ Charged Aerosol Detectors



Acclaim SEC-300 column, 5 μm, 300 x 4.6 mm Acclaim SEC-1000 column, 7 μm, 300 x 4.6 mm


Mobile phase	10 mM sodium perchlorate
Flow rate	0.35 mL/min
Injection volume	50 μL
Detection	RI
Temperature	25 °C
Analytes	(0.03% - 0.1% in mobile phase) dextran (MW 5,000,000-40,000,000), PEO (MW 895,000, 580,000, 272,000, 185,000, 80,000, 43,000, and 20,000), PEG (MW 8,300, 3,000, 1,500, 600, 400 and 200), diethylene glycol (MW 106 and ethylene glycol MW 62)



Separation of polyethylene glycols on Acclaim SEC-300 columns vs. Acclaim SEC-1000 columns

Acclaim SEC-300 column, Acclaim SEC-1000 column, 4.6 x 300 mm


Mobile phase	10 mM sodium perchlorate
Flow rate	0.35 mL/min
Injection volume	5 μL
Temperature	30 °C
Detection	RI
Sample	(5 mg/mL in mobile phase) 1. PEG MW 35,000 2. PEG MW 12,000 3. PEG MW 3,400 4. PEG MW 2,000 5. PEG MW 300



Polyacrylic acid using SEC with charged-aerosol detection

Acclaim SEC-300 column, 5 µm, 300 x 4.6 mm


Mobile phase	(A) Acetonitrile (B) Water
Flow rate	0.35 mL/min
Injection volume	35 µL
Detection	Corona III; evaporator 55 °C, Engine 40 °C, 2 Hz, filter 5, power function 1.20
Temperature	30 °C
Analyte	PAA standards in water



Separation of polyacrylic acid on Acclaim SEC-300 columns vs Acclaim SEC-1000 columns

Acclaim SEC-300 column, Acclaim SEC-1000 column, 4.6 x 300 mm


Mobile phase	10 mM sodium perchlorate
Flow rate	0.35 mL/min
Injection volume	5 µL
Temperature	30 °C
Detection	RI
Sample	Dextran, MW 10,000 (5 mg/mL in mobile phase)



Maltodextrin separation using the Acclaim SEC-1000 columns

Acclaim SEC-1000 column, 4.6 x 300 mm

Mobile phase	100 mM ammonium acetate pH 5.0
Flow rate	0.35 mL/min
Injection volume	5 µL
Temperature	25 °C
Detection	Corona <i>ultra</i> Charged Aerosol Detector
Sample	MALTRINS, 5 mg/mL each





Acclaim Mixed-Mode
columns

Acclaim Mixed-Mode HPLC columns


Mixed-Mode chromatography

Mixed-Mode chromatography provides multiple functionalities on a single chromatographic support. For example, combining both reversed-phase and ion-exchange retention mechanisms or reversed-phase and HILIC or even three mechanisms (Trinity). One major advantage of this approach is that column selectivity can easily be modified for optimal selectivity by adjusting mobile phase ionic strength, pH and/or organic solvent concentration. As a result, not only is the selectivity of a Mixed-Mode column complementary to that of reversed-phase columns, but it also allows for the development of multiple complementary selectivities on the same column under different appropriate conditions. Mixed-Mode chromatography is well-suited to retaining ionic analytes, whether hydrophobic (e.g. Naproxen) or hydrophilic (e.g. Na⁺ and Cl⁻ ions) and requires no ion-pairing agents in the method, significantly improving the MS compatibility. Most importantly, Mixed-Mode chromatography column chemistry can be customized to a desired selectivity during stationary phase design.^{1,2}

- Excellent performance: selectivity, resolution and retention
- Good for separations of active pharmaceutical ingredients (APIs), mixtures, formulations, ions
- Good when retention requirements are contradictory for a single-mode column
- Offers flexibility in method development

Bi-modal Mixed-Mode phases

The Acclaim family contains three bi-modal Mixed-Mode columns and two tri-modal columns. As shown here, bi-modal columns have both a hydrophobic arm – providing reversed-phase retention, and an ion-exchange or diol group at the tip – providing ion-exchange or HILIC retention.




Tri-modal Mixed-Mode phases


Nanopolymer Silica Hybrid technology: Tri-modal Mixed-Mode phases

Acclaim Trinity P1 and P2 columns are based on Thermo Scientific™ Nanopolymer Silica Hybrid (NSH™) technology. These are high-purity porous spherical silica particles coated with charged nanopolymer particles. The inner pores have a bi-modal functionality and the outer nanopolymers have a differing ion-exchange functionality. The spatial separation of the two ion-exchange regions allows both retention mechanisms to function simultaneously and be controlled independently.

Acclaim Trinity P1



Acclaim Trinity P2



Acclaim Mixed-Mode HPLC columns

Columns	Description
Acclaim Trinity P1	<ul style="list-style-type: none"> Reversed-phase, anion-exchange, and cation exchange functionalities For superior resolution of active pharmaceutical ingredients (API) and counter ions, as well as complex mixtures
Acclaim Trinity P2	<ul style="list-style-type: none"> HILIC, anion-exchange, and cation exchange functionalities For complex mixtures including separation pharmaceutical counterion; mono- and divalent cations and anions
Acclaim Mixed-Mode WAX-1	<ul style="list-style-type: none"> Reversed-phase and anion-exchange combined functionality For separating anionic molecules with powerful adjustable selectivity control
Acclaim Mixed-Mode WCX-1	<ul style="list-style-type: none"> Reversed-phase and cation-exchange combined functionality For separating cationic molecules with powerful adjustable selectivity control
Acclaim Mixed-Mode HILIC-1	<ul style="list-style-type: none"> Combined capability to operate in either reversed-phase or HILIC mode


Acclaim Mixed-Mode HPLC column specifications

	Acclaim Trinity P1	Acclaim Trinity P2	Acclaim Mixed-Mode WAX-1	Acclaim Mixed-Mode WCX-1	Acclaim Mixed-Mode HILIC-1
Functionality	<ul style="list-style-type: none"> Reversed-phase Weak anion exchange Strong cation exchange 	<ul style="list-style-type: none"> HILIC Strong anion exchange Weak cation exchange 	<ul style="list-style-type: none"> Reversed-phase Weak anion exchange 	<ul style="list-style-type: none"> Reversed-phase Weak cation exchange 	<ul style="list-style-type: none"> Reversed-phase HILIC
USP type	–	–	L78	L85	–
Starting material	Ultrapure silica				
Particle shape	Spherical				
Particle sizes	3 µm	3 µm	3 µm 5 µm	3 µm 5 µm	3 µm 5 µm
Average pore diameter	300 Å	300 Å	120 Å	120 Å	120 Å
Surface area	100 m ² /g	100 m ² /g	300 m ² /g	300 m ² /g	300 m ² /g
pH range	2.5–7.5	2.5–7.5	2.5–7.5	2.5–7.5	2.5–7.5

Acclaim Mixed-Mode WAX-1 columns

Designed for separating anionic molecules with powerful adjustable selectivity control

The Thermo Scientific Acclaim Mixed-Mode WAX-1 column is a novel, high-efficiency silica HPLC column that combines hydrophobic and weak anion exchange characteristics. Its unique chemistry results in a multimode separation mechanism that includes reversed-phase, anion exchange, and HILIC interactions. Selectivity can be adjusted by changing ionic strength, pH or organic solvent content.




- Adjustable selectivity
- Selectivity orthogonal to reversed-phase (RP) columns
- Ideal selectivity for anionic molecules
- Excellent column efficiency and peak asymmetry
- Multimode retention mechanisms: reversed-phase, weak anion exchange, and HILIC modes

Pain relief medicine

Acclaim Mixed-Mode WAX-1 column, 5 μm, 150 x 4.6 mm



Mobile phase	40/60 v/v Acetonitrile/buffer (6.8 g potassium monophosphate and 0.5 g pyrophosphate in 1000 g D.I. H ₂ O, pH is adjusted to 6.0 with NaOH)
Flow rate	1 mL/min
Injection volume	1 μL
Detection	UV, 220 nm
Temperature	30 °C
Analytes	1. Caffeine 2. Acetaminophen 3. Salicylamide 4. Acetyl salicylic acid (Aspirin) 5. Salicylic acid




Analysis of soft drinks

Acclaim Mixed-Mode WAX-1 column, 5 μm, 150 x 4.6 mm


Mobile phase	(A) 55/45 v/v Acetonitrile/ 0.2 M phosphate buffer, pH 6.0 (B) 57/43 v/v Acetonitrile/ 0.12 M phosphate buffer, pH 2.9
Temperature	30 °C
Flow rate	1 mL/min
Injection volume	2.5 μL
Detection	UV, 210 nm
Peaks	1. Caffeine 100 μg/ml 2. Aspartame 100 3. Acesulfame, potassium 100 4. Saccharin 100 5. Sorbate, potassium 100 6. Benzoic acid 100 7. Citrin acid 300



1 Caffeine




2. Aspartame



3. Acesulfame, potassium




4. Saccharin



5. Sorbate, potassium



6. Benzoic acid




7. Citrin acid

Acclaim Mixed-Mode WCX-1 columns

Designed for separating cationic molecules with adjustable selectivity control


Thermo Scientific Acclaim Mixed-Mode WCX-1 is a novel, high-efficiency, silica-based column, with a proprietary ligand with both hydrophobic and weak cation exchange properties. Selectivity of ionizable and neutral compounds can be controlled independently or simultaneously by tuning mobile phase ionic strength, pH or organic modifier. This column therefore can separate using multiple separation modes: reversed-phase, cation exchange, and HILIC and is recommended for a variety of industrial applications, including pharmaceutical, chemical, consumer products, foods and beverages.

- Adjustable selectivity
- Ideal selectivity for separating basic molecules
- Selectivity complementary to C18 RP columns
- Multimode separation mechanism: reversed-phase, weak cation exchange, anion-exclusion and HILIC




Pharmaceutical counterions

Acclaim Mixed-Mode WCX-1 column, 5 μm, 150 x 4.6 mm									
Mobile phase	40/60 v/v CH ₃ CN/NH ₄ OAc, pH 5.2 (20 mM total)								
Flow rate	1 mL/min								
Injection volume	5 μL								
Detection	UV, 225 nm								
Temperature	30 °C								
Analytes	<table border="0"> <tr> <td>1. Maleate 50 μg/mL</td> <td>5. Dexamethasone 60 μg/mL</td> </tr> <tr> <td>2. Ketoprofen 30 μg/mL</td> <td>6. Oxprenolol 300 μg/mL</td> </tr> <tr> <td>3. Naproxen 30 μg/mL</td> <td>7. Timolol 250 μg/mL</td> </tr> <tr> <td>4. Hydrocortisone 60 μg/mL</td> <td></td> </tr> </table>	1. Maleate 50 μg/mL	5. Dexamethasone 60 μg/mL	2. Ketoprofen 30 μg/mL	6. Oxprenolol 300 μg/mL	3. Naproxen 30 μg/mL	7. Timolol 250 μg/mL	4. Hydrocortisone 60 μg/mL	
1. Maleate 50 μg/mL	5. Dexamethasone 60 μg/mL								
2. Ketoprofen 30 μg/mL	6. Oxprenolol 300 μg/mL								
3. Naproxen 30 μg/mL	7. Timolol 250 μg/mL								
4. Hydrocortisone 60 μg/mL									



Separation of catecholamines

Acclaim Mixed-Mode WCX-1 column, 5 μm, 150 x 4.6 mm											
Mobile phase	2/98 v/v CH ₂ CN/sodium phosphate, pH 6.2 (10 mM total concentration)										
Temperature	30 °C										
Flow rate	1 mL/min										
Injection volume	5 μL										
Detection	UV, 215 nm										
Peaks	<table border="0"> <tr> <td colspan="2">(0.25 nM each)</td> </tr> <tr> <td>1. NE</td> <td></td> </tr> <tr> <td>2. E</td> <td></td> </tr> <tr> <td>3. DHBA</td> <td></td> </tr> <tr> <td>4. DA</td> <td></td> </tr> </table>	(0.25 nM each)		1. NE		2. E		3. DHBA		4. DA	
(0.25 nM each)											
1. NE											
2. E											
3. DHBA											
4. DA											




Acclaim Mixed-Mode HILIC-1

Uniquely designed for both reversed-phase and HILIC operations

The Thermo Scientific Acclaim Mixed-Mode HILIC-1 column features a unique, high-efficiency, silica-based HPLC Mixed-Mode stationary phase that combines both reversed-phase (RP) and hydrophilic interaction liquid chromatography (HILIC) properties. This combination allows both hydrophobic and hydrophilic interactions to be utilized to optimize separations.


The functional group is of a hydrophobic alkyl chain with a diol group at the terminus. This unique combination results in the adjustable selectivity, making Acclaim Mixed-Mode HILIC-1 column separate mixtures that would be impossible for a C18 column. This column is suitable for a broad range of applications, including non-ionic ethoxylated surfactants, drug metabolites, lipids, polyethylene glycols (PEGs), ethoxylated surfactants, and more.

- Can operate in both RP and HILIC modes
- Retains highly polar molecules
- Unique selectivity complementary to RP columns
- Broader application range compared with conventional diol-based columns
- High-efficiency column for high-resolution separations



Cytosine and naphthalene



Acclaim Mixed-Mode HILIC-1 column, 5 μm, 150 x 4.6 mm	
Mobile phase	CH ₃ CN/0.1 M NH ₄ OAc, pH 5.2 v/v 52/48 for RP mode v/v 92/8 for HILIC mode
Flow rate	1 mL/min
Injection volume	10 μL
Detection	UV, 254 nm
Temperature	30 °C
Analytes	1. Cytosine (100 ppm) 2. Naphthalene (100 ppm)



This surfactant separation below shows two modes of separation: on top, in the reversed-phase mode, the ethoxylated surfactant elutes as one peak, for easy total content measurements. The same column, when eluted with greater organic solvent concentration in the mobile phase (the HILIC mode), separates all the ethoxylated components individually which can be used to determine the degree of ethoxylation of the surfactant.

Separation dependency of organic solvent

Acclaim Mixed-Mode HILIC-1 column, 5 μm, 150 x 4.6 mm	
Mobile phase	(A) CH ₃ CN (B) 0.1 M NH ₄ OAc, pH 5.2
Temperature	30 °C
Flow rate	1 mL/min
Injection volume	10 μL
Detection	UV, 225 nm
Sample	IGEPAL CA-630 (0.1%)




Acclaim Trinity P1 columns

Mixed-Mode column technology combining reversed-phase, anion exchange and cation exchange functionality on a single support

The Thermo Scientific Acclaim Trinity P1 HPLC column is designed with nanopolymer silica hybrid (NSH) technology, which results in a multimode surface chemistry ideal for the simultaneous separation of drugs and their counterions. The surface chemistry concurrently provides reversed-phase, cation exchange, and anion exchange functionalities. The result is maximum flexibility in method development. Separations can be optimized easily by adjusting the chromatographic parameters (mobile phase pH, ionic strength, and organic strength).


- Ideal selectivity for simultaneous separation of API and counterion
- Adjustable selectivity by mobile phase ionic strength, electrolyte type, pH, and organic solvent
- Low bleed; compatible with MS, CAD and ELSD
- Retention of hydrophilic ionic and ionizable analytes without ion-pairing reagents
- Greater flexibility in method development: each retention mechanisms can be controlled independently



The Acclaim Trinity P1 column is able to separate both pharmaceutically related cations and anions on one column. The selectivity is ideal and peaks are symmetrical. The column is designed such that cations elute before anions. No other columns can do this separation, and in fact, this separation is part of the production qualification test.


Simultaneous separation of pharmaceutical counterions

Acclaim Trinity P1 column, 3 μ m, 100 x 3.0 mm		
Mobile phase	60/40 v/v CH ₃ CN/20 mM (total) NH ₄ OAc, pH 5	
Flow rate	0.5 mL/min	
Injection volume	2 μ L	
Detection	Corona ultra (Gain = 100 pA; Filter = med; Neb temp = 30 °C)	
Temperature	30 °C	
Analytes	(50 to 100ppm)	6. Mesylate
	1. Choline	7. Nitrate
	2. Tromethamine	8. Chloride
	3. Sodium	9. Bromide
	4. Potassium	10. Iodide
	5. Meglumine	



Simultaneous separation of pharmaceutical counterions (gradient method)


Acclaim Trinity P1 column, 3 μ m, 50 x 3.0 mm																	
Mobile phase	(A) CH ₂ N (B) DI H ₂ O (C) 0.2 M NH ₄ OAc, PH 4																
Temperature	30 °C																
Flow rate	0.5 mL/min																
Injection volume	2 μ L																
Detection	Corona <i>ultra</i> (Gain = 100 pA; Filter = med; Neb temp = 30 °C)																
Peaks	<table border="0"> <tr> <td>1. Procaine</td> <td>9. Chloride</td> </tr> <tr> <td>2. Choline</td> <td>10. Bromide</td> </tr> <tr> <td>3. Tromethamine</td> <td>11. Iodide</td> </tr> <tr> <td>4. Sodium</td> <td>12. Phosphate</td> </tr> <tr> <td>5. Potassium</td> <td>13. Malate</td> </tr> <tr> <td>6. Meglumine</td> <td>14. Tartrate</td> </tr> <tr> <td>7. Mesylate</td> <td>15. Citrate</td> </tr> <tr> <td>8. Maleate</td> <td>16. Sulfate</td> </tr> </table>	1. Procaine	9. Chloride	2. Choline	10. Bromide	3. Tromethamine	11. Iodide	4. Sodium	12. Phosphate	5. Potassium	13. Malate	6. Meglumine	14. Tartrate	7. Mesylate	15. Citrate	8. Maleate	16. Sulfate
1. Procaine	9. Chloride																
2. Choline	10. Bromide																
3. Tromethamine	11. Iodide																
4. Sodium	12. Phosphate																
5. Potassium	13. Malate																
6. Meglumine	14. Tartrate																
7. Mesylate	15. Citrate																
8. Maleate	16. Sulfate																



Gradient	-10	0	2	7	15
A%	60	60	60	10	10
B%	35	35	35	0	0
C%	5	5	5	90	90

Liquid chromatography with tandem mass spectrometry (LC-MS-MS) analysis of ethanol amines (SCX mode)

Acclaim Trinity P1 column, 3 μ m, 100 x 2.1 mm	
Chromatography conditions	
System	Dionex RSLC LC LCi system
Temperature	20 °C
Mobile phase	90% CH ₃ CN, 10% NH ₄ OAc buffer
Flow rate	600 μ L/min
Injection volume	20 μ L
Mass spectrometric conditions	
System	LC-MS-MS QTRAP
Interface	TurboSpray with Electrospray ionization
Curtain gas	15
Collision gas	Medium
IonSpray voltage	4500 V
Temperature	700 °C
Ion source gas 1	50
Ion source gas 2	20
Detection mode	Multiple reaction monitoring (MRM)




Analyte	Q1MS	Q3MS	DP	CE	CXP
EA	62	44	46	15	6
DEA	106	88	66	19	6
DEA-IS	114	78	53	24	6
MDEA	120	102	46	19	8
EDEA	134	116	51	21	8
TEA	150	132	61	19	0

Acclaim Trinity P2 column

Mixed-Mode column technology; hydrophilic interaction combining HILIC, anion exchange and cation exchange functionalities

The Thermo Scientific Acclaim Trinity P2 is a unique, high-efficiency, silica-based column specifically designed for separation of pharmaceutical counterions, including monovalent and divalent cations or anions. This column is based on nanopolymer silica hybrid (NSH) technology, which consists of high-purity porous spherical silica particles coated with charged nanopolymer particles. The inner-pore area of the silica bead is modified with a covalently bonded organic layer that provides cation-exchange retention, while the outer surface is modified with anion-exchange nano-polymer beads.




The Acclaim Trinity P2 column is aimed to complement Acclaim Trinity P1 columns to provide a total solution for pharmaceutical counter ion analysis by HPLC.

- Ideal for separating pharmaceutical counterions, including monovalent and divalent cations or anions
- Selectivity complementary to the Acclaim Trinity P1 column
- Low column bleed, compatible with CAD and MS
- Hydrolytically stable
- High-efficiency

Pharmaceutical-related anions and cations

Acclaim Trinity P2 column, 3 μ m, 100 x 3.0 mm		
Mobile phase	D.I. water and 100 mM $\text{NH}_4\text{O}^-\text{Fm}$, pH 3.65 gradient	
Flow rate	0.60 mL/min	
Injection volume	2 μ L	
Detection	Corona Veo Charged Aerosol	
Temperature	30 $^\circ\text{C}$	
Sample	0.02 – 0.10 mg/mL each in D.I.	
Analytes	1. Phosphate	7. Nitrate
	2. Sodium	8. Citrate
	3. Potassium	9. Fumarate
	4. Chloride	10. Sulfate
	5. Malate	11. Magnesium
	6. Bromide	12. Calcium

Time (min)	H_2O	0.1 M Ammonium formate, pH 3.65
-10	0.760	1.474
0	80	20
2	80	20
22	0	100
30	0	100




A broad selection of inorganic and organic ions can be used as pharmaceutical counterions. It is highly desirable to separate both pharmaceutically important anions and cations within the same analysis and in a reasonable amount of time. This figure illustrates that Acclaim Trinity P2 column provides desired selectivity for the separation of mono- and multi-valent anions and cations – a total of twelve ions.

Electrolytes in sports beverages

Acclaim Trinity P2 column, 3 µm, 50 x 3.0mm	
Mobile phase	(A) Water (B) 100 mM Ammonium formate, pH 3.65
System	Dionex UltiMate 3000 RS
Temperature	30 °C
Flow rate	0.60 mL/min
Injection volume	2 µL
Detection	Corona Veo (evaporator 55 °C, data rate 5 Hz, filter 2 sec, power function 1.5)
Samples	(A) Sports drink (Orange flavor) (B) Sports drink, zero-calorie (Fruit punch flavor)
Sample prep	Decolorized with Dionex OnGuard-II P cartridge
Peaks	<ol style="list-style-type: none"> 1. Sugars 2. Ascorbic acid 3. Phosphate 4. Sodium 5. Potassium 6. Chloride 7. Citrate 8. Acesulfame 9. Unknown 10. Magnesium 11. Calcium

Gradient	-8.0	0.0	1.0	11.0	20.0
A%	90	90	90	0	0
B%	10	10	10	100	100





Acclaim application-specific
HPLC columns

Acclaim application-specific HPLC columns

Innovative chemistries tailored for challenging and critically important applications

Application-specific columns utilize novel and unique chemistries to provide superior resolution with ease of use for key pharmaceutical, environmental and food/beverage applications. These columns are designed and tested for specific application.

Columns	Description
Acclaim Organic Acid	Fast organic acid analysis; including both aromatic and aliphatic organic acids
Acclaim Surfactant	Designed for the separation of surfactants including anionic, non-ionic, cationic, and amphoteric surfactants in various matrices
Acclaim Surfactant Plus	Designed for the separation of surfactants including anionic, non-ionic, cationic, and amphoteric surfactants in various matrices. The Surfactant Plus column is compatible with LC-MS instrumentation, and recommended for new method development
Acclaim Explosives E2	Specialty columns to comply with EPA Method 8330; baseline resolution of all 14 explosive residues targeted by the EPA
Acclaim Trinity Q1	Excellent for trace analysis of diquat and paraquat
Acclaim Carbonyl C18	Separation of DNPH derivatives of aldehydes and ketones; including regulated methods U.S. EPA 554, EPA 8315, EPA 1667, EPA TO-11, and California Air Resources Board (CARB) Method 1004
Acclaim Carbamate	Separation of carbamate pesticides specified in U.S. EPA Method 531.2
Acclaim AmG C18	Ion-pairing reversed-phase separation of aminoglycoside antibiotics

Specifications for Acclaim application-specific HPLC columns

	Acclaim Organic Acid	Acclaim Surfactant	Acclaim Surfactant Plus	Acclaim Explosives E2	Acclaim Trinity Q1	Acclaim Carbonyl C18	Acclaim Carbamate	Acclaim AmG C18
End-capped	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Particle shape	Spherical							
Particle size	3 μm 5 μm	3 μm 5 μm	3 μm 5 μm	2.2 μm 3 μm 5 μm	3 μm	2.2 μm 3 μm 5 μm	2.2 μm 3 μm 5 μm	3 μm
Average pore diameter	120 Å	120 Å	120 Å	120 Å	300 Å	120 Å	120 Å	120 Å
Surface area	300 m ² /g	300 m ² /g	300 m ² /g	300 m ² /g	100 m ² /g	300 m ² /g	300 m ² /g	300 m ² /g
pH range	2–8	2–8	2.5–7.5	2–8	2.5–7.5	2.5–7.5	2.5–7.5	0.5–10

Acclaim Organic Acid columns

Optimized and application-tested for the analysis of hydrophilic organic acids

The Thermo Scientific Acclaim Organic Acid (OA) is a silica-based reversed-phase column designed for high-efficiency, high-throughput organic acids analysis. It offers unparalleled performance for separating hydrophilic aliphatic and aromatic organic acids at low pH with UV detection.




Acclaim Acid column is the recommended column for determining small hydrophilic organic acids, C1 to C7 aliphatic acids, and hydrophilic aromatic acid and is also valuable for the analysis and quality assurance of food and beverage products, pharmaceutical preparations, plating baths, and manufacturing chemicals, chemical intermediates, and environmental samples.

- Tested to guarantee consistent hydrophilic organic acid separations
- Compatible with 100% aqueous mobile phases
- Hydrolytic stability at low-pH conditions
- Ideal selectivity for separating a wide spectrum of organic acids
- Excellent column efficiency and peak shapes for organic acids

Hydrophilic organic acids


Acclaim OA column, 5 μ m, 4 \times 250 mm

Mobile phase	100mM Na ₂ SO ₄ , pH 2.65 (adjusted with methanesulfonic acid)	
Flow rate	0.60 mL/min	
Injection volume	5 μ L	
Detection	UV, 210nm	
Temperature	30 $^{\circ}$ C	
Analytes	1. Oxalic acid 15 mg/L (ppm)	7. Acetic acid 120
	2. Tartaric acid 120	8. Citric acid 120
	3. Formic acid 180	9. Succinic acid 120
	4. Malic acid 120	10. Fumaric acid 7
	5. iso-Citric acid 120	11. cis-Aconitic acid
	6. Lactic acid 180	12. trans-Aconitic acid



Flexible methods development

Since the Acclaim OA can be operated at low pH with 100% aqueous buffers, as well as with organic solvents, many mobile phase options are available to optimize your organic acid separations. Modifying the pH of the mobile phase, as illustrated below, allows resolution and retention control.




White wine analysis

Acclaim OA column, 5 µm, 250 x 4 mm

Mobile phase	0.1 M Na ₂ SO ₄ , pH 2.68 (adjusted with MSA*)
Flow rate	0.60 mL/min
Injection volume	5 µL
Detection	UV, 210 nm
Sample	OnGuard II P
Temperature	30 °C

Peaks	1. Tartaric acid
	2. Malic acid
	3. Lactic acid
	4. Citric acid
	5. Succinic acid

*Methanesulfonic acid




Separation of seven food additives

Acclaim OA column, 3 µm, 2.1 x 150 mm

Mobile phase	(A) 100 mM KH ₂ PO ₄ pH 3 (adjusted with MSA) (B) Acetonitrile
Flow rate	0.6 mL/min
Injection volume	5 µL
Detection	UV, 210 nm
Columns temperature	50 °C (still air mode)
Pre-heater temperature	50 °C

Peaks	1. Citrate
	2. Acesulfame
	3. Saccharin
	4. Caffeine
	5. Aspartame
	6. Sorbate
	7. Benzoate



Gradient	0	5	7	7.1	10
A%	95	75	75	95	End
B%	5	26	25	5	

Acclaim Surfactant and Surfactant Plus columns

Column of choice for surfactant analysis using higher sensitivity detection: performance, versatility, throughput

The Thermo Scientific Acclaim Surfactant and Surfactant Plus columns are high-efficiency, silica-based columns designed specifically for separating a wide variety of surfactants, including anionic, cationic, nonionic, ethoxylated and amphoteric surfactants. The surface chemistry is optimized for improved performance and higher throughput.



Acclaim Surfactant column can be used with evaporative light scattering detectors (ELSD), UV-Vis detectors (UV) or refractive index (RI) detection.

The Acclaim Surfactant Plus column is similar to the original Acclaim Surfactant column and exhibits exceptionally low bleed making it ideal for use with charged aerosol detectors (CAD) and mass spectrometers (MS). The Acclaim Surfactant Plus column, like the Acclaim Surfactant columns can be used to separate a wide variety of surfactants including isomers of xylene sulfonate. Additionally, these columns can be used with suppressed conductivity detectors (SCD); non-metallic PEEK hardware is available for best compatibility with Dionex ion chromatography systems.


Surfactants are widely used in industrial, agricultural, and pharmaceutical markets, in products as diverse as pesticides, detergent powders, petroleum products, cosmetics, and pharmaceuticals. The Acclaim Surfactant and Surfactant Plus columns are designed specifically for HPLC separation of these surfactants.

- Ideal selectivity for simultaneous separation of anionic, nonionic, cationic, and amphoteric surfactants
- Surfactant Plus compatible with multiple detectors including MS, CAD, ELSD and UV
- Excellent peak shapes, especially for cationic surfactants
- Compatible with highly aqueous mobile phases
- Improved resolution for ethoxylated surfactants
- Rugged separations under a variety of conditions

Inorganic anion, hydrotropes, cationic, nonionic, amphoteric, and anionic surfactants

Acclaim Surfactant column, 5 μ m, 150 x 4.6 mm


Mobile phase	(A) CH ₃ CN (B) 0.1 M NH ₄ OAc, pH 5.4
Gradient	25% to 85% A in 25 min, then hold 85% A for 10 min
Flow rate	1 mL/min
Injection volume	25 μ L
Detection	ELS detector
Temperature	30 °C
Analytes	1. Chloride 2. Bromide 3. Nitrate 4. Xylene sulfonate 5. Laurylpyridinium chloride 6. Lauryldimethylbenzyl-ammonium chloride 7. Triton X-100 8. Cetyl betaine 9. Decyl sulfate 10. Dodecyl sulfate 11. C ₁₀ -LAS 12. C ₁₁ -LAS 13. C ₁₂ -LAS 14. C ₁₃ -LAS



Cationic surfactants

Acclaim Surfactant Plus column, 3 μ m, 150 x 3.0 mm	
Mobile phase	(A) Acetonitrile (B) 100 mM formic acid (C) Water
Flow rate	0.50 mL/min
Injection volume	5 μ L
Detection	Conductivity with blank subtraction
Temperature	25 $^{\circ}$ C
Analytes	1. Tetrabutylammonium 2. Tetrapentylammonium 3. Tetrahexylammonium 4. Tetraheptylammonium 5. Decyl-trimethylammonium 6. Dodecyl-trimethylammonium 7. Tetradecyl-trimethylammonium 8. Hexadecyl-trimethylammonium


Time (min)	%A	%B	%C
-12	5	5	90
0	5	5	90
12	40	5	55
20	40	5	55



Simultaneous analysis of cationic, nonionic, amphoteric and anionic surfactants by liquid chromatography electro spray ionization tandem mass spectrometry (LC-ESI-MS)

Acclaim Surfactant Plus column, 3 μ m, 150 x 3.0 mm	
Flow rate	0.6 mL/min
Injection volume	2 μ L
Detection	Corona <i>ultra</i> (gain = 100 pA; Filter = med; Neb temp = 20 $^{\circ}$ C)
Temperature	30 $^{\circ}$ C
Sample	Shampoo (40 x dilution with D.I. water and filtered)

Time (min)	Acetonitrile	0.05 M Ammonium Acetate, pH 5.2
-8	25	75
0	25	75
10	80	20
20	80	20





Simultaneous analysis of cationic, nonionic, amphoteric and anionic surfactants by LC-ESI-MS

Acclaim Surfactant Plus column, 3 μm, 150 x 2.1 mm	
Chromatographic conditions	
System	UltiMate 3000 RSLC
Temperature	30 °C
Mobile phase	(A) D.I. water (B) 100 mM ammonium acetate, pH 5 (C) Acetonitrile
Flow rate	0.3 mL/min

MS conditions	
System	MSQ Plus single quadrupole MS
Interface	Electrospray ionization (ESI)
Probe temp.	450 °C
Needle vol.	3 kV
Nebulizer gas	Nitrogen at 85 psi
Scan mode	Polarity switching full scan 100–1000 m/z

Time (min)	%A	%B	%C
-10	65	5	30
0	65	5	30
1	65	5	30
8	10	5	85
20	10	5	85



Acclaim AmG C18 column

Designed to provide rugged and reproducible reversed-phase chromatography of aminoglycoside antibiotics


Aminoglycoside antibiotics are commonly used as clinical and veterinary medicines to treat bacterial infections. HPLC using ion-pairing reversed-phase separations is an effective technique for simultaneous qualitative and quantitative determination of aminoglycosides.



- Rugged and reproducible
- Excellent selectivity for the HPLC of aminoglycosides
- Superior resistance to acidic conditions for long column lifetime
- Easy-to-use with only aqueous mobile phase; TFA only, or TFA/HFBA or when PFPA is needed
- Compatible with simple rugged methods; no solvents are required
- High-efficiency and throughput

Isocratic separation of gentamicin sulfate using 100 mM TFA as the mobile phase

Acclaim AmG C18 column, 3 μ m, 150 x 3.0 mm	
Mobile phase	100 mM TFA
Flow rate	0.425 mL/min
Injection volume	2 μ L
Detection	Corona Veo RS (Filter = 5.0 s; Evaporation Temp = 35 $^{\circ}$ C; Data Rate = 5 Hz; Power Function = 1.00)
Temperature	30 $^{\circ}$ C
Sample	Gentamicin (1 mg/mL)




Excellent low pH stability

The ion-pairing reversed-phase HPLC (IP-RPLC) separation of aminoglycoside antibiotics is generally performed under low pH conditions and therefore the stationary phase/column low pH stability is vital for these applications. The Acclaim AmG C18 column is specifically designed for analysis of aminoglycoside antibiotics, and compatible with these low pH conditions. The Acclaim AmG C18 columns are packed with a polymer encapsulated silica covalently bonded with C18 ligands. The polymer layer protects the siloxane linkage on the silica surface from hydrolysis when exposed to the low pH environment.

Hydrolytic stability is illustrated here, using low pH volatile perfluorinated carboxylic acids as the ion-pairing reagent, 100 mM trifluoroacetic acid (TFA) at high temperature.

Excellent low pH stability


Acclaim AmG C18 column and Typical C18 column, 3.0 × 150 mm	
Acid stress protocol	
Mobile phase	100 mM TFA
Flow rate	0.425 mL/min
Temperature	80 °C
Performance test	
Mobile phase	Acetonitrile/10 mM NH ₄ OAc, 10/90 (v/v)
Flow rate	0.425 mL/min
Injection volume	2 µL
Temperature	80 °C
Detection	UV, 220 nm
Sample	Acetanilide



Shown below is the separation of structurally related aminoglycosides include Sisomicin, netilmicin, and etimicin. Other examples (Amikacin, Kanamycin, Tobramycin, Arbekacin, Streptomycin, Ribostamycin, Paromomycin, Neomycin, Spectinomycin, Apramycin) of can be found in the product specification sheet.

Analysis of sisomicin, netilmicin, and etimicin

Acclaim AmG C18 column, 3 µm, 3.0 × 150 mm	
Mobile phase	100 mM TFA
Flow rate	0.425 mL/min
Injection volume	2 µL
Temperature	30 °C
Detection	Corona Veo RS (Filter = 5.0 s; Evaporation temp = 35 °C; Data rate = 5 Hz; Power function = 1.00)
Sample	1. Sisomicin (0.2 mg/mL) 2. Netilmicin (1 mg/mL) 3. Etimicin (1 mg/mL) 4. Mixture (0.25 mg/mL)



Acclaim Trinity Q1 columns


For trace analysis of diquat and paraquat

Thermo Scientific Acclaim Trinity Q1 columns are unique, high-efficiency, silica-based columns designed for the separation of the herbicides diquat and paraquat. These herbicides are toxic and residues are monitored in drinking water, wastewater and agricultural products. The Acclaim Trinity Q1 column is a tri-mode (WCX, WAX, RP) column based on nanopolymer silica hybrid (NSH) technology. It offers unmatched high-resolution and high-throughput trace analysis of the herbicides diquat and paraquat by LC-MS/MS and LC-UV (liquid chromatography with ultraviolet detection) methods.

- Excellent resolution of diquat and paraquat
- Good peak shape
- Fast analysis
- LC-MS compatible
- No ion-pairing reagent needed




Acclaim Trinity P1



Diquat and paraquat

Acclaim Trinity Q1 column, 3µm, 50 x 3.0mm	
Mobile phase	25% ammonium acetate (100mM, pH 5.0); 75% acetonitrile
Flow rate	0.5 mL/min
Injection volume	5 µL
Detection	Show Mass Spectrometric conditions and the scan events etc. table underneath are the peaks section
Temperature	Ambient
Mass spectrometric	Thermo Scientific™ TSQ Quantiva™ Access MAX Triple Quadrupole Mass Spectrometer
Interface	Heated Electrospary Ionization with HESI II probe
Spray voltage	1500 V
Vaporizer temp	400 °C
Sheath gas pressure	70
Aux gas pressure	10
Capillary temp	350 °C
Quantitation mode (SRM)	Selected reaction monitoring

Scan events	Precursor	Quantitative	Confirmative
		SRM (CID)	SRM (CID)
Paraquat	185	169 (27)	170 (17)
Paraquat-d ₆	193	178 (17)	
Diquat	183	157 (22)	130 (31)
Diquat-d ₃	186	158 (22)	




Method development

The Acclaim Trinity Q1 column is designed for applications using volatile buffers, such as ammonium acetate, which are compatible with MS and UV at (>225 nm). The column may be used with phosphate buffers when required. Ammonium acetate buffer is found to be effective for this application. The performance of the Acclaim Trinity Q1 column is based on reverse-phase and ion-exchange Mixed-Mode retention mechanism. The chromatography method can be optimized by adjusting mobile phase buffer concentration, solvent content, and pH.

Analysis of sisomicin, netilmicin, and etimicin


Acclaim Trinity Q1 column, 3 μ m, 3.0 \times 50 mm	
Mobile phase	75/25 v/v CH ₃ CN/ various conc. NH ₄ OAc, pH 5
Flow rate	0.60 mL/min
Injection volume	2 μ L
Temperature	30 $^{\circ}$ C
Detection	UV, 290 nm
Samples	Dq and Pq (0.1 mg/mL each)



Buffer concentration affects retentions of both diquat and paraquat. Running the separation using various buffer concentrations are shown above. Higher buffer concentration shortens retention times. If using lower buffer concentration, the retention is longer with a better the separation. Note that the resolutions were all very good for all the tested buffer concentration. For fast analysis, the 25 mM would be recommended.

Analysis of sisomicin, netilmicin, and etimicin

Acclaim Trinity Q1 column, 3 μ m, 3.0 \times 50 mm	
Mobile phase	MeCN/ 25 mM (total) NH ₄ OAc, pH 5
Flow rate	0.60 mL/min
Injection volume	2 μ L
Temperature	30 $^{\circ}$ C
Detection	UV, 290 nm
Samples	Dq and Pq (0.1 mg/mL each)



Mobile phase organic solvent content affects retention and resolution of both diquat and paraquat. At 25 mM ammonium acetate, higher acetonitrile contents give better resolution. Typically, mobile phases containing 50 to 75% acetonitrile give excellent resolution and sufficient retention times. The retention time can be adjusted depending on sample matrix and interference.

Mobile phase pH has significant effect on the resolution of diquat and paraquat. It has been determined that pH 5 \pm 0.5 is suitable pH range for this application.

Acclaim Carbamate columns

Designed for baseline separation of carbamate pesticides specified in US EPA method 531.2


Thermo Scientific Acclaim Carbamate columns are designed for baseline separation of carbamates (N-methylcarbamate and N-methylcarbamoyloxime pesticides) specified in U.S. EPA Method 531.2. Carbamate pesticides are widely used throughout the world. Drinking water and raw surface water is monitored for the presence of carbamate pesticides and related compounds applying an established U.S. EPA Method 531.2 that uses HPLC with postcolumn derivatization. LC-MS is the method of choice for the ultimate sensitivity.



- Baseline separation of carbamate pesticides specified in U.S. EPA Method 531.2
- Use with either LC/post-column derivatization/fluorescence or LC-MS detection
- Available in 2.2, 3 and 5 μm particle size
- Compatible with both binary (methanol/water) and ternary (acetonitrile/methanol/water) mobile phase gradients
- High-efficiency, extremely low column bleed, and rugged column packing

Carbamate standard – spiked rice samples

Acclaim Carbamate column, 3 μm , 150 x 3.0 mm	
Mobile phase	Methanol-H ₂ O
Gradient	Methanol, -4.0-0.0 min, 14%; 2.0 min, 20%; 8.0 min, 40%; 13.6-16 min, 70%
Flow rate	0.9 mL/min
Injection volume	250 μL
Detection	Excitation/330nm and Emission/465nm
Temperature	50 °C
Analytes	1. Aldicarb sulfoxide
	2. Aldicarb sulfone
Analytes	3. Oxamyl
	4. Methomyl
Analytes	5. 3-Hydroxy carbofuran
	6. Aldicarb
Analytes	7. Propoxur
	8. Carbofuran
Analytes	9. Carbaryl
	10. 1-Naphthol
Analytes	11. Methiocarb
	12. BDMC (I.S.)





LC-MS method

Compared to fluorescence detection, mass spectrometric detection can significantly improve method selectivity and sensitivity. Figure 2 demonstrates excellent separation, high throughput (20 min. run time), and enhanced selectivity and sensitivity with MS detection achieved on a 2.1 × 150 mm Acclaim Carbamate column. This method has been successfully applied to the determination of carbamates in various types of water samples and performance was evaluated with respect to linearity, calibration range, detection limits, and recovery from a simulated ground water matrix.⁴

DNPH aldehydes and ketones

Acclaim Carbamate column, 3 μm, 2.1 × 150 mm			
Chromatographic conditions			
System	UltiMate 3000 HPLC system		
Mobile phase	(A) Methanol (B) 1.0 mM ammonium formate (C) Water		
Flow rate	300 μL/min		
Injection volume	20 μL		
Detector	MSQ Plus single quadrupole mass spectrometer		
Mass spectrometric conditions			
Ionization interface	Electrospray ionization (ESI) positive mode		
Detection mode	Selected ion monitoring (SIM)		
Time (min)	%A	%B	%C
-4	10	5	85
0.0	10	5	85
2.0	10	5	85
15.0	65	5	30
15.1	90	5	5



Acclaim Carbonyl C18 columns

A silica-based, reversed-phase column designed specifically for separating DNPH derivatives of aldehydes and ketones

Thermo Scientific Acclaim Carbonyl C18 columns are silica-based reversed-phase columns designed specifically for separating DNPH derivatives of aldehydes and ketones. They exhibit superior resolution compared with other commercially available columns.

Aldehydes and ketones are common pollutants in air and water. Several standard methods have been developed to apply using dinitrophenylhydrazine (DNPH) to various environmental situations to measure these compounds. Some of the better known ones include California Air Resources Board (CARB) Method 1004 for vehicle exhaust, EPA Method 554 for drinking water, EPA Method 1667 for pharmaceutical wastewater, and EPA Method 8315 for general wastewater.

- Ideal selectivity for baseline resolution of DNPH derivatives of aldehydes and ketones regulated by various official methods, including EPA 554, EPA 8315, EPA 1667, EPA TO-11, and CARB 1004
- High-efficiency for UHPLC performance
- Rugged columns with good lot-to-lot reproducibility
- Proven robust methods




DNPH aldehydes and ketones

Acclaim Carbonyl RSLC column, 2.2 μm, 150 x 2.1 mm

Mobile phase	(A) D.I. water (B) Acetonitrile
Flow rate	0.400 mL/min
Injection volume	1 μL
Detection	UV, 360nm
Temperature	28 °C
Samples	Calibration mixes diluted in methanol

Analytes	1. Formaldehyde DNPH
	2. Acetaldehyde DNPH
	3. Propionaldehyde DNPH
	4. Crotonaldehyde DNPH
	5. Butyraldehyde DNPH
	6. Cyclohexanone DNPH
	7. Valeraldehyde DNPH
	8. Hexanal DNPH
	9. Heptanal DNPH
	10. Octanal DNPH
	11. Nonanal DNPH
	12. Decanal DNPH
	13. Acetone DNPH
	14. Acrolein DNPH
	15. Butanone DNPH
	16. Methacrolein DNPH
	17. Benzaldehyde DNPH
	18. Isovaleraldehyde DNPH
	19. Toluvaldehyde DNPH
	20. Xylylaldehyde DNPH




Gradient	-4.5	0.0	8.3	15.0	18.0
A%	48	48	48	0	0
B%	52	52	52	100	100

Separation of 12 carbonyl compounds regulated by EPA Method 554

Acclaim Carbonyl RSLC columns, 2.2 μm , 2.1 \times 100 mm

Mobile phase	(A) D.I. water (B) Methanol
Flow rate	0.5 mL/min
Injection volume	1 μL
Temperature	42 $^{\circ}\text{C}$
Detection	UV, 360 nm, data collection rate at 10 Hz
Sample	Calibration mix, 50 $\mu\text{g/mL}$ in methanol
Peaks	1. Formaldehyde DNPH 2. Acetaldehyde DNPH 3. Propionaldehyde DNPH 4. Crotonaldehyde DNPH 5. Butyraldehyde DNPH 6. Cyclohexanone 7. DNPH Valeraldehyde 8. Hexanal DNPH 9. Heptanal DNPH 10. Octanal DNPH 11. Nonanal DNPH 12. Decanal DNPH

Gradient	-1.7	0.0	3.4	5.5	7.0
A%	30	30	30	0	0
B%	70	70	70	100	100



References

1. Exploring Mixed-Mode Chromatography: Column Chemistry, Properties, and Applications, Xiaodong Liu, and Christopher Pohl; Thermo Fisher Scientific, Sunnyvale, CA, USA, Thermo Publication PN21137
2. http://files.alfresco.mjh.group/alfresco_images/pharma//2015/03/24/17516363-2762-4507-b9a8-6487ae731afc/PN-PITTCO-N-MixedModeChrom.pdf
3. Mixed-Mode chromatography in pharmaceutical and biopharmaceutical applications, Journal of Pharmaceutical and Biomedical Analysis, K. Zhang, X. Liu, 128 (2016) 73-88
4. Wang, L., Liu, X., Henday, S.M. and Schnute, W.C. Improved LC-MS Method for the Determination of Carbamates in Water Samples (<http://www.dionex.com/en-us/webdocs/77385-POHPLC-Carbamates-01Jul2009-LPN2295-01.pdf>).



Acclaim column selection guide

Acclaim column selection guide

Please refer to thermofisher.com/LCcolumns for more information

		Acclaim 120 C18	Acclaim 120 C8	Acclaim 300 C18	Acclaim Polar Advantage (PA)	Acclaim Polar Advantage II (PA2)	Acclaim C30	Acclaim Phenyl-1	Acclaim Trinity P1	Acclaim Trinity P2	Acclaim Mixed-Mode WAX-1	Acclaim Mixed-Mode WCX-1	Acclaim Mixed-Mode HILIC-1	Acclaim HILIC-10	Acclaim SEC-300 and SEC-1000	Acclaim Organic Acid	Acclaim Surfactant/Plus	Acclaim Carbonyl	Acclaim Explosives E2	Acclaim Carbamate	Acclaim AmG	Example Applications	
General applications	Neutral molecules	High hydrophobicity																				Fat-soluble vitamins, PAHs, glycerides	
		Intermediate hydrophobicity																					Steroids, phthalates, phenols, polyphenols
		Low hydrophobicity																					Acetaminophen, urea, polyethylene glycols
	Anionic molecules	High hydrophobicity																					NSAIDs, phospholipids
		Intermediate hydrophobicity																					Aspirin, alkyl acids, aromatic acids
		Low hydrophobicity																					Small organic acids, e.g. acetic acids
	Cationic molecules	High hydrophobicity																					Antidepressants
		Intermediate hydrophobicity																					Beta blockers, benzidines, alkaloids
		Low hydrophobicity																					Antacids, pseudoephedrine, biogenic amines
	Amphoteric/zwitterionic molecules	High hydrophobicity																					Phospholipids
		Intermediate hydrophobicity																					Amphoteric surfactants, peptides
		Low hydrophobicity																					Amino acids, aspartame, small peptides
	Mixtures of neutral, anionic, cationic molecules	Neutrals and acids																					Artificial sweeteners
		Neutrals and bases																					Cough syrup
		Acids and bases																					Drug active ingredient with counterion
Neutrals, acids, and bases																						Combination pain relievers, PEG	
Surfactants	Anionic																					SDS, LAS, laureth sulfates	
	Cationic																					Quats, benzylalkonium in medicines	
	Nonionic																					Triton X-100 in washing tank	
	Amphoteric																					Cocoamidopropyl betaine	
	Hydrotropes																					Xylenesulfonates in handsoap	
	Surfactant blends																					Nonionic and anionic surfactants	
Organic acids	Hydrophobic																					Aromatic acids, fatty acids	
	Hydrophilic																					Organic acids in soft drinks, pharmaceuticals	
Environmental contaminants	Explosives																					U.S. EPA Methods 8330, 8330B	
	Carbonyl compounds																					U.S. EPA Methods 1667, 555, OT-11; CA CARB 1004	
	Phenols																					Compounds regulated by U.S. EPA 604	
	Chlorinated/Phenoxy acids																					U.S. EPA Method 555	
	Triazines																					Compounds regulated by U.S. EPA 619	
	Nitrosamines																					Compounds regulated by U.S. EPA Method 8270	
	Benzidines																					U.S. EPA Method 605	
	Perfluorinated acids																					Dionex TN73	
	Microcystins																					ISO 20179	
	Isocyanates																					U.S. OSHA Methods 42, 47	
	Carbamate insecticides																					U.S. EPA Method 531.2	
Vitamins	Water-soluble vitamins																					Vitamins in dietary supplements	
	Fat-soluble vitamins																					Vitamin pills	
Pharmaceutical counterions	Anions																					Inorganic anions and organic acids in drugs	
	Cations																					Inorganic cations and organic bases in drugs	
	Mixture of Anions and Cations																					Screening of pharmaceutical counterions	
	API and counterions																					Naproxen Na ⁺ salt, metformin Cl ⁻ salt, etc.	
Industry	Hydrophilic cations																					Antibiotics, aminoglycosides	
	Water soluble Polymers																					PEG, PVP, PAA, PEI	



Expect reproducible results with sample prep, columns and vials



Don't see what you need? We would be happy to discuss your specific requirements. Please contact your local sales representative for custom orders.

Find out more at thermofisher.com/acclaim